

# Instructions for Use Champions® Rotary Instruments

You can find article numbers in the current Product Catalog.

## Please Note:

Please read these Instructions for Use before using the Champions® implant system. Please also follow the Champions® Basic Rules, Champions® Condensers Instructions for Use as well the Champions® Drilling and Condensing Protocols for manual and automated use, explaining drilling and Condenser sequences for insertion.

The use of the Champions® rotary instruments is restricted to dentists and doctors who are familiar with dental surgery, including diagnosis and preoperative management, considering its indication and general guidelines for dental/surgical applications as well as regulations for safety at work and prevention of accidents.

Prior to each surgical treatment, check that all required parts, instruments, and auxiliary devices are complete, functioning correctly, and available in sufficient quantity. Use the Champions® rotary instruments only if they are in good condition. Care should be taken to protect the components used inside the patient's mouth against aspiration and ingestion. Therefore, we recommend that you should undergo proper training by users with the necessary relevant experience. If in doubt regarding indication or application, refrain from usage until all items are clarified.

These Instructions for Use alone are not sufficient to ensure a professional application for treating practitioners inexperienced in Implantology. As the application of the product takes place beyond our control, we disclaim any liability for damage caused.

The treating practitioner takes sole responsibility.

## 1. Product description

Champions® rotary instruments are part of the Champions® implant systems.

- Champions (R)Evolution® implant
- Champions® one-piece implants
  - Square "New Art"
  - Square "Classic"
  - Ball-Head

Champions® rotary instruments - Drills / Insertion Aids / Mucosal Punches and Condenser are one-piece instruments.

For the preparation of the implant bed, rotary instruments are manufactured in various diameters and lengths. Champions® rotary instruments are designed for use with standard handpieces.

For this purpose, they feature a handpiece shank in accordance with EN ISO 1797 for dental connections.

Champions® Drills are equipped with laser-etched colored depth markings, providing the user with visually secured and precise drilling.

Color-coded Drill Stop Sleeves further enhance safety during depth drilling.

Champions® Condensers are instruments primarily used in soft bone for bone compaction and serve as test instruments to determine the optimal implant diameter. These are anodized in different colors according to the diameter.

Champions® Insertion Aids are intended for the placement of implants.

Champions® Mucosal Punches are suitable for preparing the gingiva.

## 1.1. Material

Designation	Material	Standard
Drills	Stainless steel 1.4197	EN ISO 7153-1
Insertion Aids/Mucosal Punches	Stainless steel 1.4197	EN ISO 7153-1
Condensers	Grade titanium 5	EN ISO 5832-3
Drill Stop Sleeves	TECAPEEK MT CLASSIX	–

## 1.2. Accessories

Champions® rotary instruments are designed for use with standard handpieces. For this purpose, they feature a handpiece shank in accordance with EN ISO 1797 for dental connections.

## 2. Delivery terms

Champions® rotary instruments are delivered in sterile packaging and bear the STERILE R symbol (gamma method). The symbol LOT refers to the batch code.

### 2.1 Cleaning, disinfection and sterilization

Champions® rotary instruments are reusable. They must be disinfected, cleaned, and sterilized after each use. Frequent reprocessing has minimal impact on Champions® rotary instruments. Their reuse has been validated for 220 applications/processing cycles.

If, due to heavy usage, significant wear or damage is visible on the Champions® rotary instruments before the completion of 220 cycles, the products must be discarded earlier.

Champions-Implants accepts no liability for clinical processing of originally sterile-delivered Champions® products by the end user. It is recommended to use the validated procedures outlined below.

Care instructions:

Automated cleaning and disinfection:

1. Place the instruments in a suitable container inside the thermos-disinfector so that the spray jet directly sprays the products.
2. Add process chemicals to the device as indicated on the product label and the thermo-disinfector manufacturer's instructions.
3. Start the Vario thermos-disinfection program, including thermal disinfection. Thermal disinfection follows A-value considerations and complies with national regulations (EN / ISO 15883).
4. After the program finishes, remove the products from the thermo-disinfector and dry them (per recommendations of the Robert-Koch-Institut (RKI [disease control/prevention]), preferably using compressed air).
5. Visually inspect for intactness and cleanliness. If you see visible contamination after the automated processing, repeat cleaning and disinfecting until you do not see any contamination. This does not apply to items that cannot be reprocessed.

## Validated process:

Step	Process
1	<b>Pre-rinse</b>  Rinse agent: Cold tap water Exposure time: 60 seconds
2	<b>Cleaning</b>  Rinse agent: Tap water Rinse temperature: 55°C Detergent: neodisher®MediClean forte / Manufacturer: Dr. WEIGERT Concentration: 0.50% Exposure time: Min, 300 seconds (RKI recommendation 600 seconds)
3	<b>Neutralizing (rinse)</b>  Rinse agent: Tap water Rinse temperature: 40°C Neutralizing agent: neodisher® N / Manufacturer: Dr. WEIGERT Concentration: 0.10% Exposure time: 60 seconds
4	<b>Final rinse</b>  Rinse agent: Tap water Rinse temperature: 40°C Exposure time: 60 seconds
5	<b>Thermal disinfection</b>  Disinfection temperature: 93°C Disinfection time: 300 seconds

## Manual cleaning and disinfection (alternative, non-validated):

1. Place the instruments in an ultrasonic device or instrument bath filled with detergents and disinfectants (lid remains closed).
2. Follow the manufacturer's instructions for concentration and exposure time for chemical disinfection in the ultrasonic device or instrument bath. The exposure time begins only after the last instrument or system component has been added to the bath and must not be shortened under any circumstances.
3. After the exposure time, thoroughly rinse the instruments with suitable water (preferably fully demineralized water [VE] to prevent residue).
4. Dry the instruments (preferably using compressed air, per RKI-recommendations).
5. Visually inspect the instruments for intactness and cleanliness. If visible contamination remains on the instruments, repeat the cleaning and chemical disinfection process until contamination is no longer visible. This does not apply to items that cannot be reprocessed. The Robert Koch-Institut (RKI) recommends automated cleaning and disinfection as the preferred method.

## Sterilization in the autoclave:

Champions® rotary instruments are sterilizable.

The components can be sterilized using steam in a vacuum process in a device compliant with EN 13060 at 134° C.

## Validated procedure:

Pre-vacuum	3 times
Sterilization temperature	134°C
Sterilization time	3 minutes
Drying time	20 minutes

For sterilization, the products must be sealed in hospital-grade packaging (paper/ pouch packages compliant with EN ISO 11607-1 and EN 868-2). Ensure that the sterilization pouches are not under tension when sealing.

In order to prevent staining and corrosion, the steam must be free of components. Recommended limits for feedwater and steam condensate components are defined by EN 13060. Do not exceed the sterilizer's maximum load capacity when sterilizing multiple instruments. Follow the device manufacturer's instructions. Do not use corroded system components anymore. Do not use damaged instruments anymore as they pose an increased risk of breakage!

#### General remarks

Observe the legal regulations concerning medical device reprocessing, valid in your country (e.g. [www.rki.de](http://www.rki.de)). The manufacturer ensures that the processing methods mentioned above are suitable for processing the concerned instruments for its reuse. The medical device operator is responsible for the processing with suitable equipment and materials by qualified staff according to the valid RKI-recommendations. For this, routine check-ups of the validated automated processing procedures are necessary. In addition, the operator must carefully evaluate the effectiveness and possible disadvantageous consequences resulting from any deviation from the procedure described here.

#### Safety and liability

Worn or damaged instruments or system components are to be discarded immediately and replaced by new ones. The above-mentioned instructions for use are to be followed. The instruments or system components shall be used only for the specified intended purpose. There will be a risk of injury if safety instructions are not followed.

#### Liability

Before using the products, users are obliged, on their own responsibility, to check the products for suitability and possibility of use for their intended purpose. Any contributory negligence by the user shall limit or entirely exclude the liability of Champions-Implants GmbH in the event of damage caused. This is especially applicable when the user has ignored Instructions for Use or safety instructions, or when he/she has accidentally misused the product.

## 2.2 Storage, use by date, packaging, and return

The product is to be stored in its original package and kept dry at room temperature and kept away from direct sunlight. Incorrect or unsafe storage can impair essential material characteristics and cause product failure.

The use by date for the first use of the product is indicated on the label. The use by date is represented by the hourglass symbol. After the expiry of the use by date indicated on the packaging, disinfect, clean, and sterilize the sterile products before use.

Only medical devices in their closed original blister packaging are sterile.

Do not perform a treatment with these components in case of an opened or damaged package, otherwise sterility and/or integrity of the products can be impaired.

These products are excluded from an exchange.

## 3. Intended purpose

Champions® rotary instruments are used to prepare and process the implant site and to insert Champions® implant systems.

## 4. Indication

Functional and aesthetic rehabilitation of the upper jaw and the lower jaw using implant systems.

## 5. Target patient group and intended user

Champions® rotary instruments are intended for use in patients with reduced or missing residual teeth. It is recommended that you should perform this treatment only in patients whose jawbone has reached maturity.

The use of Champions® rotary instruments is restricted to dental implantologists and dentists who are familiar with dental surgery, including diagnosis and preoperative management, considering its indication and general guidelines for dental/surgical applications.

Compliance with the Champions®-Basic Rules is a prerequisite for professional and safe use of the Champions® rotary instruments.

## 6. Contraindications

Risk for anatomical structures in the planned treatment site.

In addition, observe all known contraindications for dental interventions.

## 7. Warnings

- For intraoral use, Champions® products must be protected from aspiration. After accidental swallowing of products, the destination of the object is to be identified (e.g. X-rays), and necessary medical action must be undertaken.
- Champions® Basic Rules explain the use of the Champions® instruments, the drilling, Condenser, and insertion techniques, and the recommended insertion torque for transmucosal minimally invasive treatment methods.
- When preparing the implant site, take care of special structures proximities (incl. nerves, maxillary sinus, adjacent teeth). Reversible or irreversible impairment (damage) of these structures can occur.
- It is crucial that you avoid overheating and overloading of the bone (bone necrosis)..
- Ensure that the drill does not become misaligned or stuck during use (increased risk of breakage).

### 7.1 Notes

- Patients must be informed about the valid general precautions and behavioral guidelines to follow before the surgical procedure.
- To determine the exact position and depth of the drilling, it is recommended to perform computed tomography (CT) in addition to occlusal radiographs and orthopantomograms (panoramic X-rays).
- To avoid harming adjacent structures, the area surrounding the site where the instruments will be used must be carefully examined. Local anesthesia should be applied at the instrument's site of use.
- Surgical drills are subject to wear during use; timely replacement of drills is the user's responsibility. It is recommended to individually inspect the condition of each product before surgery and replace drills if necessary.
- The manufacturer reserves the right to change the design of the product, components, or its packaging, to revise Instructions for Use as well as pricing and delivery terms.
- Liability is limited to the replacement of the defective product. Further claims of any kind are excluded.
- Problems with Champions® products must be notified with details of the product (article number, batch code) to Champions-Implants GmbH, Flonheim. Serious incidents must be notified to the company Champions-Implants GmbH and to the competent statutory and regulation authorities.

## **8. Information on compatibility**

Our Champions®-implant systems offer a wide range of components for surgical and prosthetic restorations.

Please take care to use only original Champions® system connections.

You can find detailed information and indications for further system components in the current Product Catalog.

## **9. Important information for patients**

Patients should be informed about the potential side effects and complications, contraindications, warnings, and precautions associated with the treatment with Champions® products.

## **10. Disposal**

Safely dispose of contaminated or no longer usable medical devices as (clinical) health care waste in accordance with local health care directives as well as rules and regulations or directives from the state and the authorities. When separating, recycling, or disposing of packaging material, observe, where applicable, local rules and regulations from the state and authorities on packaging and packaging waste.

#### Manufacturer in the EU:

Champions-Implants GmbH  
Managing Director: Dr. med. dent. Armin Nedjat  
Im Baumfeld 30 | Champions Platz 1  
D-55237 Flonheim  
Germany  
Tel. (49) (0) 6734 - 91 40 80 | Fax (49) (0) 6734 - 10 53  
info@champions-implants.com  
champions-implants.com























CE 0297

Champions® is a registered trademark of  
Champions-Implants GmbH

Rev. 5/2024-09

#### Symbols glossary

You can find the following symbols on the product labels or on the accompanying product information.

	CE Marking with notified body reference
	Manufacturer
	Article number
	Batch code
	Date of manufacturer
	Medical device
	Non-sterile
	Sterilized using irradiation
	Use by date
	Do not resterilize
	Do not reuse
	Caution
	Consult Instructions for Use
	Do not use if package is damaged
	Temperature limit
	Keep away from sunlight
	Keep dry
	Sterile packaging
	Protective packaging with sterile barrier system inside
	Caution: U.S. law federal law restricts this device to sale by or on the order of a dental professional
	Quantity
	Max. rotation speed